

**Scharlab S.L.**

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**AC0329_21092703/1**

Product Acetonitrile, gradient 240nm/ far UV HPLC grade

Batch 21092703

AC0329

Quality release date 07/09/2020

Expiry date 09/2025

Analysis	Batch value	Specifications
Assay (G.C.)	99,99 %	≥ 99,9
Identity (IR-Spectrum)	passes test	passes test
Density (20°/4°)	0,7823 g/mL	0,779 - 0,783
Acidity	0,00019 meq/g	≤ 0,0002
Alkalinity	0,00008 meq/g	≤ 0,0001
Residue on evaporation	0,0001 %	≤ 0,0002
Water (K.F.)	0,0128 %	≤ 0,02
Gradient grade (240 nm)		
maximum background absorbance:	0,003 AU	≤ 0,01
maximum peak absorbance:	0,0003 AU	≤ 0,0015
UV Spectroscopy in a 1,0 cm cell:		
min. transmission at 200 nm	95,63 %	≥ 90
min. transmission at 205 nm	97,06 %	≥ 92
min. transmission at 210 nm	97,63 %	≥ 95
min. transmission at 220 nm	98,82 %	≥ 98
max. absorbance at 200 nm	0,019 AU	≤ 0,046
max. absorbance at 205 nm	0,013 AU	≤ 0,036
max. absorbance at 210 nm	0,010 AU	≤ 0,022
max. absorbance at 220 nm	0,005 AU	≤ 0,009

Microfiltered through membranes of pore diameter 0,22 µm.

For laboratory use only.