

**Scharlab S.L.**

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**ME0315_22002107/1**

Product Methanol, Multisolvant® HPLC grade ACS ISO UV-VIS K.F.

Batch 22002107

ME0315

Quality release date 24/08/2021

Expiry date 08/2026

Analysis	Batch value	Specifications
Assay (G.C.)	99,99 %	≥ 99,9
Identity (IR-Spectrum)	passes test	passes test
Density (20°/20°)	0,7928 g/mL	0,791 - 0,793
Appearance	clear	clear
Colour (Hazen)	1	≤ 10
Boiling point	64 - 65 ° C	64 - 65
Solubility in water	passes test	passes test
Acidity	0,00019 meq/g	≤ 0,0002
Alkalinity	0,00001 meq/g	≤ 0,0002
Chlorides (Cl)	< 0,5 ppm	≤ 0,5
Sulfates (SO4)	< 1 ppm	≤ 1
Aluminium (Al)	< 0,1 ppm	≤ 0,1
Arsenic (As)	< 0,02 ppm	≤ 0,02
Barium (Ba)	< 0,1 ppm	≤ 0,1
Beryllium (Be)	< 0,02 ppm	≤ 0,02
Bismuth (Bi)	< 0,02 ppm	≤ 0,02
Boron (B)	< 0,02 ppm	≤ 0,02
Cadmium (Cd)	< 0,05 ppm	≤ 0,05
Calcium (Ca)	< 0,5 ppm	≤ 0,5
Chromium (Cr)	< 0,02 ppm	≤ 0,02
Cobalt (Co)	< 0,02 ppm	≤ 0,02
Copper (Cu)	< 0,02 ppm	≤ 0,02
Gallium (Ga)	< 0,02 ppm	≤ 0,02
Gold (Au)	< 0,02 ppm	≤ 0,02
Indium (In)	< 0,02 ppm	≤ 0,02
Iron (Fe)	< 0,1 ppm	≤ 0,1
Lead (Pb)	< 0,1 ppm	≤ 0,1
Lithium (Li)	< 0,05 ppm	≤ 0,05
Magnesium (Mg)	< 0,1 ppm	≤ 0,1
Manganese (Mn)	< 0,02 ppm	≤ 0,02
Molybdenum (Mo)	< 0,02 ppm	≤ 0,02
Nickel (Ni)	< 0,02 ppm	≤ 0,02
Platinum (Pt) ppm	< 0,05 ppm	≤ 0,05
Silver (Ag)	< 0,02 ppm	≤ 0,02
Thallium (Tl)	< 0,02 ppm	≤ 0,02
Tin (Sn)	< 0,1 ppm	≤ 0,1
Titanium (Ti)	< 0,02 ppm	≤ 0,02
Vanadium (V)	< 0,02 ppm	≤ 0,02
Zinc (Zn)	< 0,1 ppm	≤ 0,1
Zirconium (Zr)	< 0,02 ppm	≤ 0,02
Carbonyl compounds (as CO)	< 0,001 %	≤ 0,001

Results marked with # are out of specifications.

This certificate does not exempt the user from checking the results upon receipt of the goods.
Any copy of our CoA may be obtained from our website at www.scharlab.comRicard Balladares Abella
Laboratory Technician



Scharlab S.L.

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

+34937456400

helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS

ME0315_22002107/1

Product Methanol, Multisolvant® HPLC grade ACS ISO UV-VIS K.F.

Batch 22002107

ME0315

Quality release date 24/08/2021

Expiry date 08/2026

Analysis	Batch value	Specifications
Carbonyl compounds	passes test	passes test
Substances darkened by H ₂ SO ₄	passes test	passes test
Substances reducing KMnO ₄	passes test	passes test
Residue on evaporation	0,9 ppm	≤ 2
Water (K.F.)	0,0105 %	≤ 0,03
Liquid chromatography suitability absorbance	passes test	passes test
UV Spectroscopy in a 1,0 cm cell:		
min. transmission at 207 nm	43,46 %	≥ 10
min. transmission at 210 nm	62,32 %	≥ 20
min. transmission at 220 nm	85,68 %	≥ 50
min. transmission at 225 nm	90,66 %	≥ 68
min. transmission at 230 nm	93,97 %	≥ 74
min. transmission at 240 nm	97,77 %	≥ 80
min. transmission at 250 nm	99,30 %	≥ 96
min. transmission between 260 and 400 nm	99,73 %	≥ 98
max. absorbance at 207 nm	0,362 AU	≤ 1,000
max. absorbance at 210 nm	0,205 AU	≤ 0,700
max. absorbance at 220 nm	0,067 AU	≤ 0,300
max. absorbance at 225 nm	0,043 AU	≤ 0,170
max. absorbance at 230 nm	0,027 AU	≤ 0,130
max. absorbance at 240 nm	0,010 AU	≤ 0,100
max. absorbance at 250 nm	0,003 AU	≤ 0,020
max. absorbance at 260 - 400 nm	0,001 AU	≤ 0,010

Microfiltered through membranes of pore diameter 0,22 µm.

For laboratory use only.

Results marked with # are out of specifications.
This certificate does not exempt the user from checking the results upon receipt of the goods.
Any copy of our CoA may be obtained from our website at www.scharlab.com

Ricard Balladares Abella
Laboratory Technician