

**Scharlab S.L.**

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

**CERTIFICATE OF ANALYSIS****AQ0003\_22333501/2**Product Aquagent® Complet 5, free from pyridine, one-component reagent  
for volumetric Karl Fischer titration

Batch 22333501

Quality release date 15/12/2021

Expiry date 12/2023

**AQ0003**

Analysis	Batch value	Specifications	±U
Factor	4,95	4,5 - 5,5	0,02

**Preparation**

This reagent is prepared using gravimetric and volumetric procedures.

Contains imidazole, sulfur dioxide and diethyleneglycol monoethyl ether.

**Measurement**

Batch value certified at the time of measurement.

The equipments involved in the analysis of this reagent are validated periodically.

**Storage and use**

For KF titration.

If the product is stored and unopened, this reagent is stable for 2 years from the date of manufacturing.

Air humidity will change the water content. Protect from moisture. Keep tightly closed at room temperature. The titre must be checked before use by the user.

For laboratory use only