

**Scharlab S.L.**

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**AQ0003_22624101/2**Product Aquagent® Complet 5, free from pyridine, one-component reagent
for volumetric Karl Fischer titration

Batch 22624101

Quality release date 23/03/2022

Expiry date 03/2024

AQ0003

Analysis	Batch value	Specifications	±U
Factor	4,92	4,5 - 5,5	0,01

Preparation

This reagent is prepared using gravimetric and volumetric procedures.

Contains imidazole, sulfur dioxide and diethyleneglycol monoethyl ether.

Measurement

Batch value certified at the time of measurement.

The equipments involved in the analysis of this reagent are validated periodically.

Storage and use

For KF titration.

If the product is stored and unopened, this reagent is stable for 2 years from the date of manufacturing.

Air humidity will change the water content. Protect from moisture. Keep tightly closed at room temperature. The titre must be checked before use by the user.

For laboratory use only