

**Scharlab S.L.**

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**HE0120_22604702/1**

Product Heptane, fraction from petroleum, EssentQ®

Batch 22604702

HE0120

Quality release date 17/03/2022

Expiry date 03/2027

Analysis	Batch value	Specifications
Boiling range	93 - 99 °C	93 - 99
Density (20°/4°)	0,6939 g/mL	0,680 - 0,720
Acidity	0,00001 meq/g	≤ 0,001
Copper (Cu)	< 0,2 ppm	≤ 0,2
Iron (Fe)	< 0,5 ppm	≤ 0,5
Lead (Pb)	< 0,2 ppm	≤ 0,2
Nickel (Ni)	< 0,2 ppm	≤ 0,2
Sulfur compounds (as S)	< 0,005 %	≤ 0,005
Thiophene and homologous (as C ₄ H ₄ S)	< 0,0005 %	≤ 0,0005
Residue on evaporation	0,0001 %	≤ 0,001
Water (K.F.)	0,0027 %	≤ 0,02

For laboratory use only

Results marked with # are out of specifications.

This certificate does not exempt the user from checking the results upon receipt of the goods.

Any copy of our CoA may be obtained from our website at www.scharlab.comSara Garcia Maldonado
Laboratory Technician