

**Scharlab S.L.**

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**ET0002_23782803/1**

Product Ethanol absolute, EssentQ®

Batch 23782803

ET0002

Quality release date 03/05/2023

Expiry date 05/2028

Analysis	Batch value	Specifications
Assay (G.C.) (v/v)	99,99 %	≥ 99,9
Identity (IR-Spectrum)	passes test	passes test
Density (20°/4°)	0,7896 g/mL	0,789 - 0,790
Residue on evaporation	0,0002 %	≤ 0,005
Water (v/v) (K.F.)	0,0260 %	≤ 0,1

For laboratory use only