



Scharlab S.L.

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

+34937456400

helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS

CL0347_23839509/1

Product Dichloromethane, stabilized with approx. 50 ppm of amylene,
Multisolvent® HPLC grade ACS ISO UV-VIS

Batch 23839509

Quality release date 23/05/2023

Expiry date 05/2026

CL0347

Analysis	Batch value	Specifications
Assay (G.C.)	99,99 %	≥ 99,9
Identity (IR-Spectrum)	passes test	passes test
Density (20°/4°)	1,3261 g/mL	1,323 - 1,327
Appearance	clear	clear
Colour (Hazen)	3	≤ 10
Acidity	0,00001 meq/g	≤ 0,0002
Alkalinity	< 0,0002 meq/g	≤ 0,0002
Free chlorine (as Cl)	< 0,00002 %	≤ 0,00002
Chlorides (Cl)	< 0,0001 %	≤ 0,0001
Aluminium (Al)	< 0,1 ppm	≤ 0,1
Barium (Ba)	< 0,01 ppm	≤ 0,01
Boron (B)	< 0,02 ppm	≤ 0,02
Cadmium (Cd)	< 0,01 ppm	≤ 0,01
Calcium (Ca)	< 0,3 ppm	≤ 0,3
Chromium (Cr)	< 0,02 ppm	≤ 0,02
Cobalt (Co)	< 0,02 ppm	≤ 0,02
Copper (Cu)	< 0,02 ppm	≤ 0,02
Iron (Fe)	< 0,02 ppm	≤ 0,02
Lead (Pb)	< 0,1 ppm	≤ 0,1
Magnesium (Mg)	< 0,1 ppm	≤ 0,1
Manganese (Mn)	< 0,01 ppm	≤ 0,01
Nickel (Ni)	< 0,02 ppm	≤ 0,02
Tin (Sn)	< 0,1 ppm	≤ 0,1
Zinc (Zn)	< 0,01 ppm	≤ 0,01
Carbon tetrachloride (G.C.)	< 0,001 %	≤ 0,005
Chloroform (G.C.)	< 0,001 %	≤ 0,005
Ethanol (G.C.)	< 0,005 %	≤ 0,02
Methanol (G.C.)	< 0,005 %	≤ 0,01
Formaldehyde	< 0,0005 %	≤ 0,0005
Residue on evaporation	0,0001 %	≤ 0,0002
Water (K.F.)	0,0037 %	≤ 0,01
Liquid chromatography suitability absorbance	passes test	passes test
UV Spectroscopy in a 1,0 cm cell:		
min. transmission at 235 nm	45,33 %	≥ 20
min. transmission at 240 nm	77,55 %	≥ 50
min. transmission at 245 nm	91,47 %	≥ 80
min. transmission at 248 nm	95,17 %	≥ 90
min. transmission at 255 nm	98,47 %	≥ 98
max. absorbance at 235 nm	0,344 AU	≤ 0,699
max. absorbance at 240 nm	0,110 AU	≤ 0,301

Results marked with # are out of specifications.

This certificate does not exempt the user from checking the results upon receipt of the goods.

Any copy of our CoA may be obtained from our website at www.scharlab.com

Sara Garcia Maldonado
Laboratory Technician

**Scharlab S.L.**

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**CL0347_23839509/1****Product** Dichloromethane, stabilized with approx. 50 ppm of amylene,
Multisolvent® HPLC grade ACS ISO UV-VIS**Batch** 23839509**Quality release date** 23/05/2023**Expiry date** 05/2026**CL0347**

Analysis	Batch value	Specifications
max. absorbance at 245 nm	0,039 AU	≤ 0,097
max. absorbance at 248 nm	0,021 AU	≤ 0,046
max. absorbance at 255 nm	0,007 AU	≤ 0,009

Microfiltered through membranes of pore diameter 0,22 µm.

For laboratory use only.

Results marked with # are out of specifications.
This certificate does not exempt the user from checking the results upon receipt of the goods.
Any copy of our CoA may be obtained from our website at www.scharlab.com

Sara Garcia Maldonado
Laboratory Technician