



Scharlab S.L.

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

+34937456400

helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS

ME0315_23851007/1

Product Methanol, Multisolvent® HPLC grade ACS ISO UV-VIS K.F.		Batch 23851007
ME0315		Quality release date 26/05/2023
		Expiry date 05/2028
Analysis	Batch value	Specifications
Assay (G.C.)	99,99 %	≥ 99,8
Density (20°/20°)	0,7927 g/mL	0,791 - 0,793
Colour (Hazen)	3	≤ 10
Boiling point	64 - 65 ° C	64 - 65
Solubility in water	passes test	passes test
Acidity	0,00004 meq/g	≤ 0,0003
Alkalinity	0,00001 meq/g	≤ 0,0002
Carbonyl compounds	passes test	passes test
Acetone (G.C.)	< 0,001 %	≤ 0,001
Formaldehyde (G.C.)	< 0,001 %	≤ 0,001
Acetaldehyde (G.C.)	< 0,001 %	≤ 0,001
Substances darkened by H2SO4	passes test	passes test
Substances reducing KMnO4	passes test	passes test
Absorbance test	passes test	passes test
Gradient elution test	passes test	passes test
Residue on evaporation	0,0001 %	≤ 0,001
Water (K.F.)	0,01 %	≤ 0,1
UV Spectroscopy in a 1,0 cm cell:		
min. transmission at 205 nm	27,22 %	≥ 10
min. transmission at 210 nm	58,33 %	≥ 20
min. transmission at 220 nm	81,32 %	≥ 50
min. transmission at 225 nm	87,01 %	≥ 68
min. transmission at 230 nm	91,04 %	≥ 74
min. transmission at 240 nm	96,07 %	≥ 80
min. transmission at 250 nm	98,39 %	≥ 96
min. transmission between 260 and 400 nm	98,92 %	≥ 98
max. absorbance at 205 nm	0,565 AU	≤ 1,000
max. absorbance at 210 nm	0,234 AU	≤ 0,700
max. absorbance at 220 nm	0,090 AU	≤ 0,300
max. absorbance at 225 nm	0,060 AU	≤ 0,170
max. absorbance at 230 nm	0,041 AU	≤ 0,130
max. absorbance at 240 nm	0,017 AU	≤ 0,100
max. absorbance at 250 nm	0,007 AU	≤ 0,020
max. absorbance at 260 - 400 nm	0,005 AU	≤ 0,010

ACS, ISO, reag. Ph.Eur. R1, reag. Ph.Eur. R2.

Microfiltered through membranes of pore diameter 0,22 µm.

For laboratory use only.

Results marked with # are out of specifications.
This certificate does not exempt the user from checking the results upon receipt of the goods.
Any copy of our CoA may be obtained from our website at www.scharlab.com

Marta Martinez Casals
Laboratory Technician



Scharlab S.L.

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS

ME0315_23851007/1

Results marked with # are out of specifications.
This certificate does not exempt the user from checking the results upon receipt of the goods.
Any copy of our CoA may be obtained from our website at www.scharlab.com

Marta Martinez Casals
Laboratory Technician

A handwritten signature in black ink, appearing to be "Marta", written over a horizontal line.