

**Scharlab S.L.**

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**ME0315_25674607/1**

Product	Methanol, Multisolvant® HPLC grade ACS ISO UV-VIS K.F.	Batch	25674607
ME0315		Quality release date	28/03/2025
		Expiry date	03/2030

Analysis	Batch value	Specifications
Assay (G.C.)	99,99 %	≥ 99,8
Density (20°/20°)	0,7928 g/mL	0,791 - 0,793
Colour (Hazen)	1	≤ 10
Boiling point	64 - 65 ° C	64 - 65
Solubility in water	passes test	passes test
Acidity	0,00013 meq/g	≤ 0,0003
Alkalinity	0,00001 meq/g	≤ 0,0002
Carbonyl compounds	passes test	passes test
Acetone (G.C.)	< 0,001 %	≤ 0,001
Formaldehyde (G.C.)	< 0,001 %	≤ 0,001
Acetaldehyde (G.C.)	< 0,001 %	≤ 0,001
Substances darkened by H2SO4	passes test	passes test
Substances reducing KMnO4	passes test	passes test
Absorbance test	passes test	passes test
Gradient elution test	passes test	passes test
Residue on evaporation	0,0001 %	≤ 0,001
Water (K.F.)	0,02 %	≤ 0,1
UV Spectroscopy in a 1,0 cm cell:		
min. transmission at 205 nm	32,71 %	≥ 10
min. transmission at 210 nm	61,78 %	≥ 20
min. transmission at 220 nm	82,94 %	≥ 50
min. transmission at 225 nm	88,13 %	≥ 68
min. transmission at 230 nm	91,98 %	≥ 74
min. transmission at 240 nm	96,59 %	≥ 80
min. transmission at 250 nm	98,68 %	≥ 96
min. transmission between 260 and 400 nm	99,45 %	≥ 98
max. absorbance at 205 nm	0,485 AU	≤ 1,000
max. absorbance at 210 nm	0,209 AU	≤ 0,700
max. absorbance at 220 nm	0,081 AU	≤ 0,300
max. absorbance at 225 nm	0,055 AU	≤ 0,170
max. absorbance at 230 nm	0,036 AU	≤ 0,130
max. absorbance at 240 nm	0,015 AU	≤ 0,100
max. absorbance at 250 nm	0,006 AU	≤ 0,020
max. absorbance at 260 - 400 nm	0,002 AU	≤ 0,010

ACS, ISO, reag. Ph.Eur. R1, reag. Ph.Eur. R2.

Microfiltered through membranes of pore diameter 0,22 µm

For laboratory use only

Results marked with # are out of specifications.
This certificate does not exempt the user from checking the results upon receipt of the goods.
Any copy of our CoA may be obtained from our website at www.scharlab.com

Marta Martínez Casals
Laboratory Technician



Scharlab S.L.

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS

ME0315_25674607/1

A handwritten signature in black ink, appearing to read "Marta", written over a horizontal line.