

**Scharlab S.L.**

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**ME0315_26078607/1**

Product	Methanol, Multisolvant® HPLC grade ACS ISO UV-VIS K.F.	Batch	26078607
ME0315		Quality release date	18/09/2025
		Expiry date	09/2030

Analysis	Batch value	Specifications
Assay (G.C.)	99,99 %	≥ 99,8
Density (20°/20°)	0,7926 g/mL	0,791 - 0,793
Colour (Hazen)	2	≤ 10
Boiling point	64 - 65 ° C	64 - 65
Solubility in water	passes test	passes test
Acidity	0,00013 meq/g	≤ 0,0003
Alkalinity	0,00001 meq/g	≤ 0,0002
Carbonyl compounds	passes test	passes test
Acetone (G.C.)	< 0,001 %	≤ 0,001
Formaldehyde (G.C.)	< 0,001 %	≤ 0,001
Acetaldehyde (G.C.)	< 0,001 %	≤ 0,001
Substances darkened by H2SO4	passes test	passes test
Substances reducing KMnO4	passes test	passes test
Absorbance test	passes test	passes test
Gradient elution test	passes test	passes test
Residue on evaporation	0,0001 %	≤ 0,001
Water (K.F.)	0,0164 %	≤ 0,1
UV Spectroscopy in a 1,0 cm cell:		
min. transmission at 205 nm	36,46 %	≥ 10
min. transmission at 210 nm	69,80 %	≥ 20
min. transmission at 220 nm	89,13 %	≥ 50
min. transmission at 225 nm	93,69 %	≥ 68
min. transmission at 230 nm	95,19 %	≥ 74
min. transmission at 240 nm	98,01 %	≥ 80
min. transmission at 250 nm	99,15 %	≥ 96
min. transmission between 260 and 400 nm	99,53 %	≥ 98
max. absorbance at 205 nm	0,438 AU	≤ 1,000
max. absorbance at 210 nm	0,156 AU	≤ 0,700
max. absorbance at 220 nm	0,050 AU	≤ 0,300
max. absorbance at 225 nm	0,028 AU	≤ 0,170
max. absorbance at 230 nm	0,021 AU	≤ 0,130
max. absorbance at 240 nm	0,009 AU	≤ 0,100
max. absorbance at 250 nm	0,004 AU	≤ 0,020
max. absorbance at 260 - 400 nm	0,002 AU	≤ 0,010

ACS, ISO, reag. Ph.Eur. R1, reag. Ph.Eur. R2.

Microfiltered through membranes of pore diameter 0,22 µm

For laboratory use only

Results marked with # are out of specifications.
This certificate does not exempt the user from checking the results upon receipt of the goods.
Any copy of our CoA may be obtained from our website at www.scharlab.com

Marta Martínez Casals
Laboratory Technician



Scharlab S.L.

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS

ME0315_26078607/1

Results marked with # are out of specifications.
This certificate does not exempt the user from checking the results upon receipt of the goods.
Any copy of our CoA may be obtained from our website at www.scharlab.com

Marta Martínez Casals
Laboratory Technician

A handwritten signature in black ink, appearing to be "Marta", written over a horizontal line.