



## Certificate of Analysis

**Product Name :** Water Analytical Grade  
Ph.Eur.  
**Reference :** WATR-00P-25K  
**Mol. formula:** H<sub>2</sub>O  
**Mol. Weight :** 18,02  
**Batch no. :** 18320.36.071  
**Manufacturing date :** 07/2015  
**Shelf life :** 08/2017  
**Analyzed on :** 30/07/2015  
**Use :** For lab use only. For professional use only.

Sr no.	Tests	Specifications	Results
1	Total organic carbon:	≤ 0.5 mg/L	≤ 0.1 mg/L
2	Conductivity (20 °C):	≤ 4.3 µS/cm	≤ 4.3 µS/cm
3	Nitrates:	≤ 0.2 ppm	≤ 0.2 ppm
4	Heavy metals :	≤ 0.1 ppm	≤ 0.1 ppm
5	Acidity or alkalinity:	Passes test	Passes test
6	Oxidisable substances:	Passes test	Passes test
7	Chlorides:	Passes test	Passes test
8	Ammonium:	≤ 0.2 ppm	≤ 0.2 ppm
9	Calcium and magnesium	Passes test	Passes test
10	Residue on evaporation:	≤ 0.001 %	≤ 0.001 %
11	Microbial contamination (TAMC):	≤ 100 CFU/ml	≤ 100 CFU/ml

This above product complies as per the specifications of labkem

Quality Department, date : 05/08/2015

This document has been produced electronically and it is valid without signature



a business unit of labbox labware s.l

Remences, 102 - 08304 Mataró (Barcelona) Spain - Phone +34 937 532 084 - Fax +34 937 909 532  
info@labkem.com - www.labkem.com