

Certificate of Analysis

Agua Extra Pure Grade **Product Name:**

Ph.Eur

WATR-00P-20K Reference:

Mol. formula: H_2O Mol. Weight: 18,02

> Batch no.: 18988.37.906A

Manufacturing date: 11/2015

Shelf life: 11/2017 Analyzed on : 20/11/2015

> Use: For lab use only. For professional use only.

Sr no.	Tests	Specifications	Results
1	Total organic carbon:	≤ 0.5 mg/L	≤ 0.1 mg/L
2	Conductivity (20 ℃):	≤ 4.3 µS/cm	≤ 4.3 µS/cm
3	Nitrates:	≤ 0.2 ppm	≤ 0.2 ppm
4	Heavy metals :	≤ 0.1 ppm	≤ 0.1 ppm
5	Acidity or alkalinity:	Passes test	Passes test
6	Oxidisable substances:	Passes test	Passes test
7	Chlorides:	Passes test	Passes test
8	Ammonium:	≤ 0.2 ppm	≤ 0.2 ppm
9	Calcium and magnesium	Passes test	Passes test
10	Residue on evaporation:	≤ 0.001 %	≤ 0.001 %
11	Microbial contamination (TAMC):	≤ 100 CFU/mI	≤ 100 CFU/ml

This above product complies as per the specifications of labkem

Quality Department, date: 27/01/2016



