

Certificate of Analysis

European Pharmacopoeia

0.1 M Sodium Methoxide

Product No:	3007100
Lot No:	371016F2
Expiry date:	28/06/2017

Mean measured value:	0.0998 M @ 20°C
Date of measurement:	23/06/2016

Specification:
0.0995 - 0.1005M

Method:
Test in accordance with in-house acidmetric titration.

Traceability:
This volumetric solution was checked by means of Hydrochloric acid Analytical Volumetric Standard. This volumetric standard is directly traceable to a Standard Reference Material of National Institute of Standards and Technology (USA), 723e Tris(hydroxymethyl)aminomethane. All raw materials used to prepare this product are of high purity.

Products are manufactured under an NSAI registered I.S EN ISO9001:2008 Quality System, registration no: 19.2769

Date of issue of the certificate: 23/06/2016

Rubán McDonagh

QC Technician Ruban Mc Donagh

Complementary information relative to this product is available at www.reagecon.com
Page 1 of 1. This Certificate must not be reproduced except in full. Rev-16F8

Reagecon Diagnostics Ltd.
Shannon Free Zone, Shannon, Co. Clare, Ireland.
Tel +353 61 472622, Fax: +353 61 472642
Email: sales@reagecon.ie, www.reagecon.com