

Scharlab S.L.

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa 08181 Sentmenat

\(+34937456400

CERTIFICATE OF ANALYSIS

Product	Sodium hydroxide, solution 0,05 mol/l (0,05 N)		Batch	19790701	
SO0453			Quality release date	06/05/2019	
			Expiry date	05/2022	
Analysis		Batch value	Specifications	5	±U
Factor		1,000	0,999 - 1,001		0,001

Preparation

This volumetric solution is prepared using gravimetric and volumetric procedures. 1ml= 0,0020g NaOH

Traceability

This solution was checked using a Scharlau's potassium hydrogen phthalate volumetric standard. This Scharlau's volumetric standard is traceable to Standard Reference Material from NIST SRM 84L potassium hydrogen phthalate.

Uncertainty

It characterises the dispersion of the values that could be attributed to the mesurand. The limits of the uncertainty are given at a confidence level of 95% (k=2).

Measurement

The batch value is determined by means of potentiometric method (alkalimetric) at 20°C.

The equipments involved in the analysis of this solution are validated periodically.

Batch value certified at the time of measurement.

Storage and use

For volumetric titration.

If the product is stored and unopened, this solution is stable for 3 years from the date of manufacturing.

Keep tightly closed at room temperature. Avoid exposure to light and moisture.

We suggest shaking the bottle before use in order to homogenize the content.

The titre should be checked periodically by the user.

Ju _